Hydrogen Sulfide

Sensoric H2S 2E 50 S



Sensoric H2S 2E 50 S

FEATURES

Amperometric 2 electrode sensor cell fast response high reliability high selectivity

TYPICAL APPLICATIONS

TLV-monitoring, leakage detection portable applications
Oil & Petrochemical industry, water treatment plants

PART NUMBER INFORMATION

MINI	0141-124-30009
SENSORIC CLASSIC	0141-124-30069
CTL4 series adaptation	0141-124-30049
CTL 7 series adaptation	0141-124-30079



Sensoric H2S 2E 50 S

TECHNICAL SPECIFICATIONS

Measuring Range 0–50 ppm

Sensitivity Range 380 nA/ppm ± 80 nA/ ppm

Zero Current at $20\,^{\circ}$ C $< \pm 200\,^{\circ}$ A Resolution at $20\,^{\circ}$ C $< 0.7\,^{\circ}$ ppm not required Linearity < 5% full scale

Response Time at 20 ℃

< 15 s calculated from 2 min. exposure time
 < 30 s calculated from 2 min. exposure time

Long Term Sensitivity Drift < 10% per 6 months

Operation Conditions

Temperature Range -20 °C to +40 °C

Humidity Range 15–90% r.H., non–condensing


Effect of Humidity an abrupt change of r. H. will cause a short term drift in zero reading of <1 ppm

Sensor Life Expectancy > 48 months Warranty 24 months



Sensoric H2S 2E 50 S


OUTPUT vs. TEMPERATURE:





Sensoric H2S 2E 50 S

ZERO READING vs. TEMPERATURE:





Sensoric H2S 2E 50 S

CROSS SENSITIVITIES AT 20 ℃

Gas	Concentration	Reading [ppm]
Carbon Monoxide	100 ppm	<1
Chlorine	1 ppm	<±1
Nitrogen Dioxide	10 ppm	- 2
Sulfur Dioxide	20 ppm	3
Hydrogen	10000 ppm	<10
Isopropanol	200 ppm	0
Methane	2.2 %	0
Nitric Oxide	100 ppm	<±1

Notes:

- 1. Interference factors may differ from sensor to sensor and with life time. It is not adviseable to calibrate with interference gases.
- 2. This table does not claim to be complete. The sensor might also be sensitive to other gases.



Safety Note

This sensor is designed to be used in safety critical applications. To ensure that the sensor and/or instrument in which it is used, are operating properly, it is a requirement that the function of the device is confirmed by exposure to target gas (bump check) before each use of the sensor and/or instrument. Failure to carry out such tests may jeopardize the safety of people and property.

Attention

Use of the Sensoric range sensors requires complete understanding of the instructions. Before using Sensoric range sensors please carefully read 'Application Notes' which can be found at www.citytech.com under the heading 'Support' -> 'Application Notes' -> 'Sensoric'

Product Safety Data Sheets (PSDS) can be obtained at www.citytech.com under the heading 'Support' -> 'Product Safety Datasheets'

For further assistance on sensor selection and use, please contact a member of the Technical Sales team.

